

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
 Data File : VW021965.D
 Acq On : 17 Jan 2022 10:34
 Operator : SY/VA
 Sample : VW0117SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK446

Quant Time: Jan 18 00:37:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 18 00:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	135249	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	137965	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	76077	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	57543	22.891	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.560%
7) Chloroethane-d5	2.885	69	39548	17.422	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.680%
11) 1,1-Dichloroethene-d2	4.019	63	76399	19.250	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.000%
21) 2-Butanone-d5	7.086	46	23250	65.527	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.060%
24) Chloroform-d	7.653	84	112057	25.646	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.600%
26) 1,2-Dichloroethane-d4	8.305	65	59331	28.119	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.480%
32) Benzene-d6	8.275	84	229675	25.656	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.640%
36) 1,2-Dichloropropane-d6	9.274	67	60836	25.783	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.120%
41) Toluene-d8	10.323	98	226185	25.064	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.240%
43) trans-1,3-Dichloroprop...	10.573	79	26116	26.069	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	104.280%
47) 2-Hexanone-d5	10.927	63	20494	64.009	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.020%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	49381	26.764	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.040%
66) 1,2-Dichlorobenzene-d4	13.853	152	87965	28.120	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
Data File : VW021965.D
Acq On : 17 Jan 2022 10:34
Operator : SY/VA
Sample : VW0117SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK446

Quant Time: Jan 18 00:37:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 18 00:36:09 2022
Response via : Initial Calibration

