

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
 Data File : VW021967.D
 Acq On : 17 Jan 2022 12:07
 Operator : SY/VA
 Sample : VIBLK593
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK593

Quant Time: Jan 18 00:37:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 18 00:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	142909	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	143435	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	83568	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	51460	19.374	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.480%
7) Chloroethane-d5	2.885	69	34777	14.499	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.000%
11) 1,1-Dichloroethene-d2	4.019	63	67931	16.199	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.800%
21) 2-Butanone-d5	7.080	46	18378	49.020	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.040%
24) Chloroform-d	7.653	84	95709	20.731	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.920%
26) 1,2-Dichloroethane-d4	8.305	65	48999	21.978	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.920%
32) Benzene-d6	8.275	84	197078	21.175	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.680%
36) 1,2-Dichloropropane-d6	9.274	67	52496	21.400	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.600%
41) Toluene-d8	10.323	98	196169	20.909	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.640%
43) trans-1,3-Dichloroprop...	10.579	79	20891	20.058	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.240%
47) 2-Hexanone-d5	10.921	63	16147	48.509	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.020%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	40552	21.140	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.560%
66) 1,2-Dichlorobenzene-d4	13.853	152	73295	21.330	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
Data File : VW021967.D
Acq On : 17 Jan 2022 12:07
Operator : SY/VA
Sample : VIBLK593
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK593

Quant Time: Jan 18 00:37:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 18 00:36:09 2022
Response via : Initial Calibration

