

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
 Data File : VW021983.D
 Acq On : 17 Jan 2022 18:35
 Operator : SY/VA
 Sample : N1168-14
 Misc : 6.40g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLX8

Quant Time: Jan 18 00:41:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 18 00:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	159759	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.627	117	161640	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	93698	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	43322	14.590	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.360%
7) Chloroethane-d5	2.879	69	35944	13.405	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.600%
11) 1,1-Dichloroethene-d2	4.019	63	57702	12.309	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	49.240%
21) 2-Butanone-d5	7.086	46	22305	53.219	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.440%
24) Chloroform-d	7.646	84	99497	19.278	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.120%
26) 1,2-Dichloroethane-d4	8.305	65	54564	21.893	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.560%
32) Benzene-d6	8.274	84	190650	18.177	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	72.720%
36) 1,2-Dichloropropane-d6	9.274	67	55060	19.917	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.680%
41) Toluene-d8	10.323	98	182410	17.252	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	69.000%
43) trans-1,3-Dichloroprop...	10.579	79	21521	18.336	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.360%
47) 2-Hexanone-d5	10.926	63	19277	51.390	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.780%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	50350	23.292	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.160%
66) 1,2-Dichlorobenzene-d4	13.853	152	78522	20.381	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
Data File : VW021983.D
Acq On : 17 Jan 2022 18:35
Operator : SY/VA
Sample : N1168-14
Misc : 6.40g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLX8

Quant Time: Jan 18 00:41:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 18 00:36:09 2022
Response via : Initial Calibration

