

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
 Data File : VW021985.D
 Acq On : 17 Jan 2022 19:23
 Operator : SY/VA
 Sample : N1155-10
 Misc : 6.99g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLW4

Quant Time: Jan 18 00:41:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 18 00:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	156076	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	153472	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	89140	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	40642	14.010	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.040%
7) Chloroethane-d5	2.879	69	32410	12.372	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	49.480%
11) 1,1-Dichloroethene-d2	4.013	63	54409	11.880	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	47.520%
21) 2-Butanone-d5	7.080	46	26999	65.939	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.880%
24) Chloroform-d	7.653	84	96851	19.208	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.840%
26) 1,2-Dichloroethane-d4	8.305	65	54351	22.322	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.280%
32) Benzene-d6	8.275	84	187993	18.878	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.520%
36) 1,2-Dichloropropane-d6	9.268	67	53748	20.477	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.920%
41) Toluene-d8	10.323	98	179089	17.840	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	71.360%
43) trans-1,3-Dichloroprop...	10.579	79	21692	19.465	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.840%
47) 2-Hexanone-d5	10.921	63	22715	63.778	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	127.560%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	52734	25.693	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	75447	20.584	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	82.320%
Target Compounds						
					Qvalue	
13) Acetone	4.129	43	12357	36.529	ug/L	85
22) 2-Butanone	7.183	43	3630	8.395	ug/L	78
33) Benzene	8.323	78	10726	1.140	ug/L	100
51) Chlorobenzene	11.652	112	15819	2.123	ug/L	89
52) Ethylbenzene	11.731	91	515664	43.980	ug/L	95
53) m,p-Xylene	11.841	106	237039	48.347	ug/L	89
54) o-Xylene	12.164	106	22743	4.997	ug/L	89
60) Isopropylbenzene	12.463	105	603375	49.707	ug/L	93
62) 1,3,5-Trimethylbenzene	12.938	105	76849	7.645	ug/L	92
63) 1,2,4-Trimethylbenzene	13.249	105	517642	52.215	ug/L	93
67) 1,2-Dichlorobenzene	13.865	146	8027	1.502	ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
Data File : VW021985.D
Acq On : 17 Jan 2022 19:23
Operator : SY/VA
Sample : N1155-10
Misc : 6.99g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLW4

Quant Time: Jan 18 00:41:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 18 00:36:09 2022
Response via : Initial Calibration









#13
Acetone
Concen: 36.529 ug/L
RT: 4.129 min Scan# 30
Delta R.T. -0.012 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23




Instrument :
MSVOA_W
ClientSampleId :
BGLW4


Tgt Ion: 43 Resp: 12357
Ion Ratio Lower Upper
43 100
58 33.9 0.0 52.8



#22
2-Butanone
Concen: 8.395 ug/L
RT: 7.183 min Scan# 866
Delta R.T. 0.006 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Tgt Ion: 43 Resp: 3630
Ion Ratio Lower Upper
43 100
72 34.2 11.8 35.4









#33
Benzene
Concen: 1.140 ug/L
RT: 8.323 min Scan# 1053
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Instrument : MSVOA_W
ClientSampleId : BGLW4




Tgt Ion: 78 Resp: 10726




#51
Chlorobenzene
Concen: 2.123 ug/L
RT: 11.652 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Tgt Ion: 112 Resp: 15819

Ion	Ratio	Lower	Upper
112	100		
114	35.1	22.4	41.6
77	52.6	50.2	75.4









#52
Ethylbenzene
Concen: 43.980 ug/L
RT: 11.731 min Scan# 1
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23




Instrument :
MSVOA_W
ClientSampleId :
BGLW4


Tgt Ion: 91 Resp: 515664
Ion Ratio Lower Upper
91 100
106 33.3 21.5 39.9



#53
m,p-Xylene
Concen: 48.347 ug/L
RT: 11.841 min Scan# 1630
Delta R.T. 0.006 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Tgt Ion:106 Resp: 237039
Ion Ratio Lower Upper
106 100
91 196.7 149.9 278.3









#54
o-Xylene
Concen: 4.997 ug/L
RT: 12.164 min Scan# 1
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23




Instrument :
MSVOA_W
ClientSampleId :
BGLW4


Tgt Ion:106 Resp: 22743
Ion Ratio Lower Upper
106 100
91 202.2 153.6 285.3



#60
Isopropylbenzene
Concen: 49.707 ug/L
RT: 12.463 min Scan# 1732
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Tgt Ion:105 Resp: 603375
Ion Ratio Lower Upper
105 100
120 28.8 20.1 30.1
77 14.5 13.8 20.8









#62
1,3,5-Trimethylbenzene
Concen: 7.645 ug/L
RT: 12.938 min Scan# 11
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23




Instrument :
MSVOA_W
ClientSampleId :
BGLW4


Tgt Ion:105 Resp: 76849
Ion Ratio Lower Upper
105 100
120 52.2 37.4 56.2



#63
1,2,4-Trimethylbenzene
Concen: 52.215 ug/L
RT: 13.249 min Scan# 1861
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Tgt Ion:105 Resp: 517642
Ion Ratio Lower Upper
105 100
120 48.0 35.0 52.6





#67
1,2-Dichlorobenzene
Concen: 1.502 ug/L
RT: 13.865 min Scan# 1962
Delta R.T. -0.000 min
Lab File: VW021985.D
Acq: 17 Jan 2022 19:23

Instrument :
MSVOA_W
ClientSampleId :
BGLW4

Tgt Ion:	146	Resp:	8027
Ion	Ratio	Lower	Upper
146	100		
111	49.4	33.1	49.7
148	66.0	47.4	71.2

