

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011822\
 Data File : VW021990.D
 Acq On : 18 Jan 2022 11:00
 Operator : SY/VA
 Sample : N1129-06RE
 Misc : 6.13g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBLZ9RE

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/19/2022
 Supervised By :Mahesh Dadoda 01/19/2022

Quant Time: Jan 19 00:31:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jan 19 00:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	9306	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.633	117	7538	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	2382	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	2900	16.766	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	67.080%		
7) Chloroethane-d5	2.891	69	2318	14.841	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	59.360%		
11) 1,1-Dichloroethene-d2	4.037	63	4125	15.106	ug/L	0.02
Spiked Amount 25.000	Range 45 - 110		Recovery =	60.440%		
21) 2-Butanone-d5	7.098	46	898m	36.783	ug/L	0.02
Spiked Amount 50.000	Range 20 - 135		Recovery =	73.560%		
24) Chloroform-d	7.652	84	6402	21.295	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	85.160%		
26) 1,2-Dichloroethane-d4	8.305	65	3457	23.812	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	95.240%		
32) Benzene-d6	8.280	84	12222	24.988	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	99.960%		
36) 1,2-Dichloropropane-d6	9.274	67	3524	27.335	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	109.320%		
41) Toluene-d8	10.323	98	9806	19.888	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	79.560%		
43) trans-1,3-Dichloroprop...	10.579	79	761m	13.903	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	55.600%		
47) 2-Hexanone-d5	10.926	63	183	10.461	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	20.920%		
56) 1,1,2,2-Tetrachloroeth...	12.694	84	1874	18.589	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	74.360%		
66) 1,2-Dichlorobenzene-d4	13.853	152	1706	17.418	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	69.680%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011822\
Data File : VW021990.D
Acq On : 18 Jan 2022 11:00
Operator : SY/VA
Sample : N1129-06RE
Misc : 6.13g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBLZ9RE

Quant Time: Jan 19 00:31:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jan 19 00:30:45 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/19/2022
Supervised By : Mahesh Dadoda 01/19/2022

