

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011922\
 Data File : VW022001.D
 Acq On : 19 Jan 2022 11:55
 Operator : SY/VA
 Sample : N1168-11
 Misc : 5.06g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jan 20 00:24:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jan 20 00:22:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	112789	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	111755	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	62508	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	43632	20.813	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.240%
7) Chloroethane-d5	2.885	69	32946	17.404	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	69.600%
11) 1,1-Dichloroethene-d2	4.019	63	58394	17.644	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.560%
21) 2-Butanone-d5	7.092	46	19004	64.226	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.460%
24) Chloroform-d	7.653	84	87807	24.098	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	8.311	65	46357	26.345	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	8.275	84	176418	24.328	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.320%
36) 1,2-Dichloropropane-d6	9.274	67	48341	25.292	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.160%
41) Toluene-d8	10.323	98	169733	23.219	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.880%
43) trans-1,3-Dichloroprop...	10.579	79	18718	23.066	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.280%
47) 2-Hexanone-d5	10.926	63	16918	65.233	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	130.460%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	40055	26.800	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	63992	24.897	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011922\
Data File : VW022001.D
Acq On : 19 Jan 2022 11:55
Operator : SY/VA
Sample : N1168-11
Misc : 5.06g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jan 20 00:24:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jan 20 00:22:54 2022
Response via : Initial Calibration

