

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011922\
 Data File : VW022002.D
 Acq On : 19 Jan 2022 12:19
 Operator : SY/VA
 Sample : N1114-21
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jan 20 00:24:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jan 20 00:22:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	119562	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	121804	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	71519	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	44453	20.004	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.000%
7) Chloroethane-d5	2.885	69	32088	15.990	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.960%
11) 1,1-Dichloroethene-d2	4.019	63	57673	16.439	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.760%
21) 2-Butanone-d5	7.086	46	17494	55.773	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.540%
24) Chloroform-d	7.653	84	89589	23.194	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.760%
26) 1,2-Dichloroethane-d4	8.311	65	46828	25.106	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.440%
32) Benzene-d6	8.275	84	176233	22.298	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.200%
36) 1,2-Dichloropropane-d6	9.274	67	48518	23.291	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.160%
41) Toluene-d8	10.323	98	168959	21.206	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.840%
43) trans-1,3-Dichloroprop...	10.579	79	18860	21.324	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.280%
47) 2-Hexanone-d5	10.927	63	15660	55.401	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.800%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	40078	24.603	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	66428	22.589	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011922\
Data File : VW022002.D
Acq On : 19 Jan 2022 12:19
Operator : SY/VA
Sample : N1114-21
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jan 20 00:24:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jan 20 00:22:54 2022
Response via : Initial Calibration

