

Data Path : W:\HPCHEM1\MSVOA W\DATA\VW012418\
 Data File : VW000947.D
 Acq On : 24 Jan 2018 14:51
 Operator : JC/SY
 Sample : J1187-08RE
 Misc : 6.32G/5.0ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-106-4A(70-72)RE

Quant Time: Jan 24 18:03:15 2018
 Quant Method : W:\HPCHEM1\MSVOA_W\METHOD\82W012318S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 23 13:31:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	36518	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	55431	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	46616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	17468	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	28932	82.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	165.32%	
35) Dibromofluoromethane	7.88	113	11935	35.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.30%	
50) Toluene-d8	10.33	98	64861	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.63	95	17900	36.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.44%	
Target Compounds						
16) Acetone	4.13	43	5828	53.39	ug/l	91
20) Methylene Chloride	4.92	84	1408	1.56	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA W\DATA\VW012418\
Data File : VW000947.D
Acq On : 24 Jan 2018 14:51
Operator : JC/SY
Sample : J1187-08RE
Misc : 6.32G/5.0ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-106-4A(70-72)RE

Quant Time: Jan 24 18:03:15 2018
Quant Method : W:\HPCHEM1\MSVOA_W\METHOD\82W012318S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 23 13:31:00 2018
Response via : Initial Calibration

