

Data Path : W:\HPCHEM1\MSVOA W\DATA\VW012418\
 Data File : VW000957.D
 Acq On : 24 Jan 2018 19:28
 Operator : JC/SY
 Sample : J1241-05
 Misc : 6.34G/5.0ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EPE-106-4A(110-112)

Quant Time: Jan 25 05:56:41 2018
 Quant Method : W:\HPCHEM1\MSVOA_W\METHOD\82W012318S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 23 13:31:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	232568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	352652	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	325483	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	169093	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	125354	56.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.46%	
35) Dibromofluoromethane	7.88	113	94949	43.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.90%	
50) Toluene-d8	10.33	98	421802	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	12.63	95	149252	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
Target Compounds						
16) Acetone	4.12	43	8491	26.05	ug/l	# 87
20) Methylene Chloride	4.92	84	39233	17.60	ug/l	96
30) Chloroform	7.68	83	5706	1.50	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA W\DATA\VW012418\
Data File : VW000957.D
Acq On : 24 Jan 2018 19:28
Operator : JC/SY
Sample : J1241-05
Misc : 6.34G/5.0ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EPE-106-4A(110-112)

Quant Time: Jan 25 05:56:41 2018
Quant Method : W:\HPCHEM1\MSVOA_W\METHOD\82W012318S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 23 13:31:00 2018
Response via : Initial Calibration

