

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012624\
 Data File : VW027857.D
 Acq On : 26 Jan 2024 10:57
 Operator : SY/MD
 Sample : VW0126SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK420

Quant Time: Jan 26 23:37:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Jan 25 00:24:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	318300	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	289197	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	148929	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	76169	19.606	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.440%
7) Chloroethane-d5	2.899	69	63339	18.597	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.400%
11) 1,1-Dichloroethene-d2	4.027	65	33639	19.249	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.000%
21) 2-Butanone-d5	7.081	46	34936	54.383	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.760%
24) Chloroform-d	7.648	84	173285	22.514	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.040%
26) 1,2-Dichloroethane-d4	8.307	65	84656	24.277	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.120%
32) Benzene-d6	8.276	84	368073	22.425	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.680%
36) 1,2-Dichloropropane-d6	9.270	67	103235	22.734	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.920%
41) Toluene-d8	10.319	98	337122	21.970	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.880%
43) trans-1,3-Dichloroprop...	10.575	79	42521	24.090	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.360%
47) 2-Hexanone-d5	10.922	63	33454	56.126	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.260%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	82596	25.533	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.120%
66) 1,2-Dichlorobenzene-d4	13.848	152	118067	22.552	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012624\
Data File : VW027857.D
Acq On : 26 Jan 2024 10:57
Operator : SY/MD
Sample : VW0126SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK420

Quant Time: Jan 26 23:37:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Jan 25 00:24:22 2024
Response via : Initial Calibration

