

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012725\
 Data File : VW031599.D
 Acq On : 27 Jan 2025 17:42
 Operator : SY/MD
 Sample : Q1190-10
 Misc : 4.91g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A44S9

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 01/28/2025
 Supervised By : Mahesh Dadoda 01/28/2025

Quant Time: Jan 28 00:27:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011025SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 28 00:17:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	722738	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	488749	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	136649	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	257140	22.673	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	90.680%		
7) Chloroethane-d5	2.893	69	194395	24.002	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	96.000%		
11) 1,1-Dichloroethene-d2	4.015	65	112455	22.286	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	89.160%		
21) 2-Butanone-d5	7.081	46	176462	96.793	ug/L	-0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	193.580%#		
24) Chloroform-d	7.648	84	504842	26.990	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	107.960%		
26) 1,2-Dichloroethane-d4	8.301	65	292622	31.785	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	127.160%		
32) Benzene-d6	8.270	84	999977	32.572	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	130.280%		
36) 1,2-Dichloropropane-d6	9.270	67	324617	35.812	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	143.240%#		
41) Toluene-d8	10.319	98	752376	27.148	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	108.600%		
43) trans-1,3-Dichloroprop...	10.575	79	113481	30.487	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	121.960%		
47) 2-Hexanone-d5	10.922	63	117460	126.980	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	253.960%#		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	235345	42.456	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	169.840%#		
66) 1,2-Dichlorobenzene-d4	13.848	152	129427	27.004	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	108.000%		
Target Compounds						
13) Acetone	4.149	43	88947m	63.393	ug/L	
22) 2-Butanone	7.185	43	17286	7.214	ug/L	98
33) Benzene	8.319	78	74726	2.234	ug/L	100
46) Tetrachloroethene	10.861	164	99952	15.488	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW012725\
Data File : VW031599.D
Acq On : 27 Jan 2025 17:42
Operator : SY/MD
Sample : Q1190-10
Misc : 4.91g/10mL/MSVOA_W/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A44S9

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/28/2025
Supervised By : Mahesh Dadoda 01/28/2025

Quant Time: Jan 28 00:27:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011025SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 28 00:17:29 2025
Response via : Initial Calibration

