

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013024\
 Data File : VW027863.D
 Acq On : 30 Jan 2024 13:32
 Operator : SY/MD
 Sample : P1286-01
 Misc : 4.75g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH6S2

Quant Time: Jan 30 22:01:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 30 22:00:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	550845	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	465212	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	199549	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	119377	17.756	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.040%
7) Chloroethane-d5	2.899	69	71865	12.192	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	48.760%
11) 1,1-Dichloroethene-d2	4.021	65	53609	17.726	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.920%
21) 2-Butanone-d5	7.087	46	46623	41.937	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.880%
24) Chloroform-d	7.648	84	260501	19.557	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.240%
26) 1,2-Dichloroethane-d4	8.307	65	119377	19.782	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	79.120%
32) Benzene-d6	8.276	84	547309	20.729	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.920%
36) 1,2-Dichloropropane-d6	9.270	67	152640	20.896	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.600%
41) Toluene-d8	10.319	98	514507	20.844	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.360%
43) trans-1,3-Dichloroprop...	10.575	79	57739	20.335	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.320%
47) 2-Hexanone-d5	10.922	63	42161	43.971	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.940%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	110023	21.143	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.560%
66) 1,2-Dichlorobenzene-d4	13.848	152	150441	21.446	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013024\
Data File : VW027863.D
Acq On : 30 Jan 2024 13:32
Operator : SY/MD
Sample : P1286-01
Misc : 4.75g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH6S2

Quant Time: Jan 30 22:01:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 30 22:00:20 2024
Response via : Initial Calibration

