

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013124\
 Data File : VW027880.D
 Acq On : 31 Jan 2024 15:38
 Operator : SY/MD
 Sample : P1293-09
 Misc : 4.54g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH931

Quant Time: Feb 01 00:37:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 00:34:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	545180	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	501307	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	229578	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	113750	17.095	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.360%
7) Chloroethane-d5	2.899	69	96103	16.474	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.880%
11) 1,1-Dichloroethene-d2	4.021	65	53346	17.822	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.280%
21) 2-Butanone-d5	7.087	46	59720	54.275	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.560%
24) Chloroform-d	7.648	84	283660	21.517	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.080%
26) 1,2-Dichloroethane-d4	8.307	65	141444	23.682	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.720%
32) Benzene-d6	8.276	84	589494	20.719	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.880%
36) 1,2-Dichloropropane-d6	9.270	67	172851	21.959	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.840%
41) Toluene-d8	10.319	98	567983	21.353	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.400%
43) trans-1,3-Dichloroprop...	10.575	79	70316	22.981	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.920%
47) 2-Hexanone-d5	10.922	63	57992	56.127	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.260%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	147150	26.242	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.960%
66) 1,2-Dichlorobenzene-d4	13.848	152	192718	23.880	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013124\
Data File : VW027880.D
Acq On : 31 Jan 2024 15:38
Operator : SY/MD
Sample : P1293-09
Misc : 4.54g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH931

Quant Time: Feb 01 00:37:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 01 00:34:16 2024
Response via : Initial Calibration

