

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013124\
 Data File : VW027871.D
 Acq On : 31 Jan 2024 12:05
 Operator : SY/MD
 Sample : P1293-02
 Misc : 5.53g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH924

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/01/2024
 Supervised By :Mahesh Dadoda 02/06/2024

Quant Time: Feb 01 00:35:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 00:34:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	527931	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	469399	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	240563	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	114167	17.718	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	70.880%		
7) Chloroethane-d5	2.899	69	100274	17.751	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	71.000%		
11) 1,1-Dichloroethene-d2	4.021	65	51635	17.814	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	71.240%		
21) 2-Butanone-d5	7.087	46	56452	52.982	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	105.960%		
24) Chloroform-d	7.648	84	275682	21.595	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	86.400%		
26) 1,2-Dichloroethane-d4	8.307	65	132643m	22.934	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	91.720%		
32) Benzene-d6	8.276	84	559878	21.015	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	84.080%		
36) 1,2-Dichloropropane-d6	9.276	67	163498	22.183	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	88.720%		
41) Toluene-d8	10.319	98	534278	21.452	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	85.800%		
43) trans-1,3-Dichloroprop...	10.575	79	65634	22.909	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	91.640%		
47) 2-Hexanone-d5	10.922	63	52298	54.057	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	108.120%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	134609	25.637	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	102.560%		
66) 1,2-Dichlorobenzene-d4	13.848	152	200758	23.740	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	94.960%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013124\
Data File : VW027871.D
Acq On : 31 Jan 2024 12:05
Operator : SY/MD
Sample : P1293-02
Misc : 5.53g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH924

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/01/2024
Supervised By : Mahesh Dadoda 02/06/2024

Quant Time: Feb 01 00:35:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 01 00:34:16 2024
Response via : Initial Calibration

