

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013124\
 Data File : VW027879.D
 Acq On : 31 Jan 2024 15:15
 Operator : SY/MD
 Sample : P1293-08
 Misc : 6.04g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH930

Quant Time: Feb 01 00:37:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 00:34:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	549204	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	489636	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	241621	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	106343	15.865	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.440%
7) Chloroethane-d5	2.899	69	69079	11.755	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	47.000%
11) 1,1-Dichloroethene-d2	4.021	65	47367	15.709	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.840%
21) 2-Butanone-d5	7.087	46	54886	49.517	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.040%
24) Chloroform-d	7.654	84	255332	19.226	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.920%
26) 1,2-Dichloroethane-d4	8.307	65	124117	20.629	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.520%
32) Benzene-d6	8.276	84	530195	19.079	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.320%
36) 1,2-Dichloropropane-d6	9.276	67	151608	19.720	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.880%
41) Toluene-d8	10.319	98	498351	19.182	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.720%
43) trans-1,3-Dichloroprop...	10.575	79	61068	20.434	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.720%
47) 2-Hexanone-d5	10.922	63	50287	49.830	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.660%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	130033	23.742	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.960%
66) 1,2-Dichlorobenzene-d4	13.848	152	169947	20.008	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW013124\
Data File : VW027879.D
Acq On : 31 Jan 2024 15:15
Operator : SY/MD
Sample : P1293-08
Misc : 6.04g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH930

Quant Time: Feb 01 00:37:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 01 00:34:16 2024
Response via : Initial Calibration

