

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
 Data File : VW027900.D
 Acq On : 01 Feb 2024 13:05
 Operator : SY/MD
 Sample : P1303-08
 Misc : 5.35g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0FH8

Quant Time: Feb 01 21:39:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 21:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	311176	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	261588	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	113588	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	67375	17.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.960%
7) Chloroethane-d5	2.899	69	57316	17.214	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.840%
11) 1,1-Dichloroethene-d2	4.021	65	30195	17.674	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.680%
21) 2-Butanone-d5	7.093	46	25016	39.832	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.660%
24) Chloroform-d	7.648	84	156015	20.734	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.920%
26) 1,2-Dichloroethane-d4	8.307	65	72922	21.391	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.560%
32) Benzene-d6	8.276	84	314887	21.209	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.840%
36) 1,2-Dichloropropane-d6	9.270	67	91712	22.328	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.320%
41) Toluene-d8	10.325	98	298015	21.471	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.880%
43) trans-1,3-Dichloroprop...	10.575	79	29971	18.772	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.080%
47) 2-Hexanone-d5	10.922	63	21443	39.772	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.540%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	70839	24.210	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.840%
66) 1,2-Dichlorobenzene-d4	13.848	152	89064	22.305	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.240%
Target Compounds						Qvalue
16) Methylene chloride	4.929	84	14909	3.305	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
Data File : VW027900.D
Acq On : 01 Feb 2024 13:05
Operator : SY/MD
Sample : P1303-08
Misc : 5.35g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0FH8

Quant Time: Feb 01 21:39:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 01 21:37:27 2024
Response via : Initial Calibration

