

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
 Data File : VW027894.D
 Acq On : 01 Feb 2024 10:43
 Operator : SY/MD
 Sample : P1293-05
 Misc : 5.31g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH927

Quant Time: Feb 01 21:38:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 21:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	581948	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	516028	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	247439	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	118375	16.666	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.680%
7) Chloroethane-d5	2.899	69	97725	15.694	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.760%
11) 1,1-Dichloroethene-d2	4.021	65	54566	17.078	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.320%
21) 2-Butanone-d5	7.088	46	45488	38.729	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.460%
24) Chloroform-d	7.648	84	277403	19.713	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.840%
26) 1,2-Dichloroethane-d4	8.307	65	127037	19.926	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	79.720%
32) Benzene-d6	8.276	84	576785	19.694	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.760%
36) 1,2-Dichloropropane-d6	9.270	67	161022	19.873	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.480%
41) Toluene-d8	10.319	98	547385	19.992	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.960%
43) trans-1,3-Dichloroprop...	10.575	79	62012	19.689	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.760%
47) 2-Hexanone-d5	10.922	63	42353	39.822	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.640%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	119014	20.619	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.480%
66) 1,2-Dichlorobenzene-d4	13.849	152	192657	22.149	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
Data File : VW027894.D
Acq On : 01 Feb 2024 10:43
Operator : SY/MD
Sample : P1293-05
Misc : 5.31g/10mL/MSVOA_W/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH927

Quant Time: Feb 01 21:38:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 01 21:37:27 2024
Response via : Initial Calibration

