

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
 Data File : VW027910.D
 Acq On : 01 Feb 2024 17:04
 Operator : SY/MD
 Sample : P1303-17
 Misc : 5.68g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 C0FH6

Quant Time: Feb 01 21:41:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 21:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	485214	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	444049	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	213691	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	94772	16.003	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	64.000%		
7) Chloroethane-d5	2.899	69	78553	15.130	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	60.520%		
11) 1,1-Dichloroethene-d2	4.021	65	44831	16.829	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	67.320%		
21) 2-Butanone-d5	7.081	46	46161	47.137	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	94.280%		
24) Chloroform-d	7.648	84	240189	20.471	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	81.880%		
26) 1,2-Dichloroethane-d4	8.307	65	114535	21.547	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	86.200%		
32) Benzene-d6	8.276	84	500520	19.860	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	79.440%		
36) 1,2-Dichloropropane-d6	9.276	67	141734	20.328	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	81.320%		
41) Toluene-d8	10.318	98	473318	20.089	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	80.360%		
43) trans-1,3-Dichloroprop...	10.575	79	55194	20.365	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	81.440%		
47) 2-Hexanone-d5	10.922	63	44926	49.088	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	98.180%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	116650	23.485	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	93.920%		
66) 1,2-Dichlorobenzene-d4	13.848	152	165557	22.039	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	88.160%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
 Data File : VW027910.D
 Acq On : 01 Feb 2024 17:04
 Operator : SY/MD
 Sample : P1303-17
 Misc : 5.68g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0FH6

Quant Time: Feb 01 21:41:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 21:37:27 2024
 Response via : Initial Calibration

