

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
 Data File : VW027915.D
 Acq On : 01 Feb 2024 19:04
 Operator : SY/MD
 Sample : P1303-22
 Misc : 5.45g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0FK3

Quant Time: Feb 01 21:42:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 01 21:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	320247	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	289308	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	138911	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	52702	13.483	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.920%
7) Chloroethane-d5	2.893	69	45010	13.135	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.520%
11) 1,1-Dichloroethene-d2	4.021	65	24675	14.034	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.120%
21) 2-Butanone-d5	7.075	46	29511	45.659	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.320%
24) Chloroform-d	7.648	84	136065	17.571	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	70.280%
26) 1,2-Dichloroethane-d4	8.307	65	67635	19.278	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.120%
32) Benzene-d6	8.270	84	275522	16.780	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	67.120%
36) 1,2-Dichloropropane-d6	9.270	67	80859	17.800	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	71.200%
41) Toluene-d8	10.319	98	259416	16.899	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.600%
43) trans-1,3-Dichloroprop...	10.575	79	31235	17.689	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	70.760%
47) 2-Hexanone-d5	10.922	63	27379	45.916	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.840%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	72497	22.402	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	13.849	152	96195	19.699	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020124\
Data File : VW027915.D
Acq On : 01 Feb 2024 19:04
Operator : SY/MD
Sample : P1303-22
Misc : 5.45g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0FK3

Quant Time: Feb 01 21:42:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 01 21:37:27 2024
Response via : Initial Calibration

