

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
 Data File : VW027927.D
 Acq On : 02 Feb 2024 14:08
 Operator : SY/MD
 Sample : P1286-10
 Misc : 5.80g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH6X1

Quant Time: Feb 02 23:03:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Feb 02 23:01:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	113160	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	102847	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	44210	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	23532	17.038	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.160%
7) Chloroethane-d5	2.899	69	18477	15.260	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.040%
11) 1,1-Dichloroethene-d2	4.021	65	9901	15.936	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.760%
21) 2-Butanone-d5	7.081	46	12889	56.435	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.880%
24) Chloroform-d	7.648	84	54402	19.881	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.520%
26) 1,2-Dichloroethane-d4	8.307	65	28524	23.009	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.040%
32) Benzene-d6	8.276	84	107734	18.456	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.840%
36) 1,2-Dichloropropane-d6	9.270	67	32393	20.059	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.240%
41) Toluene-d8	10.325	98	93863	17.200	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.800%
43) trans-1,3-Dichloroprop...	10.575	79	11712	18.658	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.640%
47) 2-Hexanone-d5	10.922	63	10127	47.775	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.540%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	28980	25.191	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	33807	21.753	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
Data File : VW027927.D
Acq On : 02 Feb 2024 14:08
Operator : SY/MD
Sample : P1286-10
Misc : 5.80g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH6X1

Quant Time: Feb 02 23:03:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 02 23:01:12 2024
Response via : Initial Calibration

