

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
 Data File : VW027921.D
 Acq On : 02 Feb 2024 11:26
 Operator : SY/MD
 Sample : P1293-11
 Misc : 5.09g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/05/2024
 Supervised By :Mahesh Dadoda 02/06/2024

Quant Time: Feb 02 23:02:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Feb 02 23:01:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	3681	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	3120	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	1304	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.357	65	915	20.366	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.480%
7) Chloroethane-d5	2.899	69	808	20.514	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.040%
11) 1,1-Dichloroethene-d2	4.033	65	337m	16.675	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.680%
21) 2-Butanone-d5	0.000	46	0d	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
24) Chloroform-d	7.648	84	2190	24.604	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.400%
26) 1,2-Dichloroethane-d4	8.301	65	1180	29.261	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	117.040%
32) Benzene-d6	8.270	84	4202	23.729	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.920%
36) 1,2-Dichloropropane-d6	9.270	67	1361	27.781	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.120%
41) Toluene-d8	10.325	98	3982	24.054	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.200%
43) trans-1,3-Dichloroprop...	10.581	79	209	10.975	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	43.920%
47) 2-Hexanone-d5	0.000	63	0	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	926	26.533	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.120%
66) 1,2-Dichlorobenzene-d4	13.848	152	1002	21.859	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
Data File : VW027921.D
Acq On : 02 Feb 2024 11:26
Operator : SY/MD
Sample : P1293-11
Misc : 5.09g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/05/2024
Supervised By : Mahesh Dadoda 02/06/2024

Quant Time: Feb 02 23:02:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 02 23:01:12 2024
Response via : Initial Calibration

