

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
 Data File : VW027931.D
 Acq On : 02 Feb 2024 15:44
 Operator : SY/MD
 Sample : P1286-12
 Misc : 4.80g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH6S5

Quant Time: Feb 02 23:04:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Feb 02 23:01:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	438330	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	391150	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	193022	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	81192	15.176	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.720%
7) Chloroethane-d5	2.899	69	70978	15.133	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.520%
11) 1,1-Dichloroethene-d2	4.021	65	36361	15.109	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.440%
21) 2-Butanone-d5	7.087	46	45686	51.642	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.280%
24) Chloroform-d	7.648	84	210279	19.839	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.360%
26) 1,2-Dichloroethane-d4	8.307	65	102309	21.305	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.240%
32) Benzene-d6	8.276	84	416097	18.743	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.960%
36) 1,2-Dichloropropane-d6	9.270	67	122687	19.976	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.920%
41) Toluene-d8	10.319	98	396419	19.100	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.400%
43) trans-1,3-Dichloroprop...	10.575	79	47928	20.076	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.320%
47) 2-Hexanone-d5	10.922	63	44054	54.645	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.280%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	107896	24.660	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.640%
66) 1,2-Dichlorobenzene-d4	13.848	152	144642	21.317	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
Data File : VW027931.D
Acq On : 02 Feb 2024 15:44
Operator : SY/MD
Sample : P1286-12
Misc : 4.80g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH6S5

Quant Time: Feb 02 23:04:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 02 23:01:12 2024
Response via : Initial Calibration

