

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
 Data File : VW027932.D
 Acq On : 02 Feb 2024 16:08
 Operator : SY/MD
 Sample : P1286-15
 Misc : 5.56g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH6S6

Quant Time: Feb 02 23:04:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Feb 02 23:01:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	453671	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	400514	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	182194	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	85997	15.531	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.120%
7) Chloroethane-d5	2.899	69	76813	15.823	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.280%
11) 1,1-Dichloroethene-d2	4.021	65	39778	15.970	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.880%
21) 2-Butanone-d5	7.081	46	45036	49.186	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.380%
24) Chloroform-d	7.648	84	227022	20.694	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.760%
26) 1,2-Dichloroethane-d4	8.307	65	109047	21.940	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.760%
32) Benzene-d6	8.270	84	459055	20.195	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.760%
36) 1,2-Dichloropropane-d6	9.270	67	134206	21.340	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.360%
41) Toluene-d8	10.318	98	430285	20.248	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.000%
43) trans-1,3-Dichloroprop...	10.575	79	51817	21.197	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.800%
47) 2-Hexanone-d5	10.922	63	43547	52.753	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.500%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	111721	24.937	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	144789	22.607	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020224\
Data File : VW027932.D
Acq On : 02 Feb 2024 16:08
Operator : SY/MD
Sample : P1286-15
Misc : 5.56g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH6S6

Quant Time: Feb 02 23:04:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM010824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 02 23:01:12 2024
Response via : Initial Calibration

