

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020323\
 Data File : VW025272.D
 Acq On : 03 Feb 2023 12:30
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK406

Quant Time: Feb 03 13:30:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Feb 03 11:28:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	412237	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	365894	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	142193	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	91693	22.986	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.960%
7) Chloroethane-d5	2.904	69	80468	22.718	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.880%
11) 1,1-Dichloroethene-d2	4.019	63	180834	16.348	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.400%
21) 2-Butanone-d5	7.080	46	46821	45.212	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.420%
24) Chloroform-d	7.647	84	240727	22.279	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.120%
26) 1,2-Dichloroethane-d4	8.305	65	127047	24.139	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.560%
32) Benzene-d6	8.275	84	494976	22.320	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.280%
36) 1,2-Dichloropropane-d6	9.274	67	141495	21.807	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.240%
41) Toluene-d8	10.323	98	449300	22.666	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.680%
43) trans-1,3-Dichloroprop...	10.579	79	57559	23.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.960%
47) 2-Hexanone-d5	10.921	63	39084	53.724	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.440%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	84528	19.156	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	76.640%
66) 1,2-Dichlorobenzene-d4	13.847	152	145470	29.487	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	117.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020323\
Data File : VW025272.D
Acq On : 03 Feb 2023 12:30
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK406

Quant Time: Feb 03 13:30:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020323SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 03 11:28:47 2023
Response via : Initial Calibration

