

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020524\
 Data File : VW027960.D
 Acq On : 05 Feb 2024 17:43
 Operator : SY/MD
 Sample : P1352-08
 Misc : 5.60g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0FN9

Quant Time: Feb 06 04:08:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 06 04:06:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	517155	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	439971	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	188143	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	100385	16.573	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.280%
7) Chloroethane-d5	2.893	69	100513	21.878	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.520%
11) 1,1-Dichloroethene-d2	4.021	65	50916	19.646	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.600%
21) 2-Butanone-d5	7.081	46	56198	59.713	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.420%
24) Chloroform-d	7.648	84	288304	23.842	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.360%
26) 1,2-Dichloroethane-d4	8.307	65	145566	26.654	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	8.270	84	558871	23.742	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.960%
36) 1,2-Dichloropropane-d6	9.270	67	156667	24.547	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.200%
41) Toluene-d8	10.319	98	533571	23.690	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.760%
43) trans-1,3-Dichloroprop...	10.575	79	64182	24.453	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.800%
47) 2-Hexanone-d5	10.922	63	55780	64.252	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.500%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	126525	26.523	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.080%
66) 1,2-Dichlorobenzene-d4	13.848	152	144807	22.644	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.560%
Target Compounds						
16) Methylene chloride	4.923	84	20452	3.164	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020524\
Data File : VW027960.D
Acq On : 05 Feb 2024 17:43
Operator : SY/MD
Sample : P1352-08
Misc : 5.60g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0FN9

Quant Time: Feb 06 04:08:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 06 04:06:06 2024
Response via : Initial Calibration

