

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW020620\
 Data File : VW014892.D
 Acq On : 06 Feb 2020 11:52
 Operator : SY/VA
 Sample : L1324-19
 Misc : 4.99G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAT74

Manual Integrations
 APPROVED

MMDadoda
 2/10/2020 12:33:14 PM

Quant Time: Feb 07 06:14:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 07 05:58:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	948842	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	845148	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	307896	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	302919	19.28	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.12%
7) Chloroethane-d5	2.89	69	293113	23.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.24%
10) 1,1-Dichloroethene-d2	4.01	63	457498	15.94	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.76%
20) 2-Butanone-d5	7.08	46	145248	32.95	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	65.90%
24) Chloroform-d	7.65	84	635079m	23.57	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.28%
26) 1,2-Dichloroethane-d4	8.31	65	327434	21.96	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.84%
29) Benzene-d6	8.27	84	1275490	23.47	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.88%
33) 1,2-Dichloropropane-d6	9.27	67	427103	24.12	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.48%
37) Toluene-d8	10.32	98	1084274	22.67	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.68%
38) trans-1,3-Dichloropropene-	10.57	79	129202	18.47	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.88%
39) 2-Hexanone-d5	10.93	63	101098	31.32	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.64%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	248464	18.63	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	74.52%
61) 1,2-Dichlorobenzene-d4	13.85	152	280401	23.48	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.92%

Target Compounds					Ovalue
13) Acetone	4.13	43	15876	3.489 ug/L	74
44) Toluene	10.38	91	44916	0.761 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020620\
Data File : VW014892.D
Acq On : 06 Feb 2020 11:52
Operator : SY/VA
Sample : L1324-19
Misc : 4.99G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAT74

Manual Integrations
APPROVED

MMDadoda
2/10/2020 12:33:14 PM

Quant Time: Feb 07 06:14:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 07 05:58:53 2020
Response via : Initial Calibration

