

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020623\
 Data File : VW025296.D
 Acq On : 06 Feb 2023 15:20
 Operator : SY/MD
 Sample : 01422-15
 Misc : 4.99g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK02

Quant Time: Feb 07 00:20:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 07 00:17:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	254922	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	226643	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	113484	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.367	65	50712	20.557	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.240%
7) Chloroethane-d5	2.897	69	42509	19.407	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.640%
11) 1,1-Dichloroethene-d2	4.019	63	109799	16.052	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.200%
21) 2-Butanone-d5	7.080	46	33944	53.005	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.020%
24) Chloroform-d	7.647	84	153562	22.983	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.920%
26) 1,2-Dichloroethane-d4	8.305	65	88581	27.217	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.880%
32) Benzene-d6	8.275	84	297447	21.654	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	9.274	67	81354	20.242	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.960%
41) Toluene-d8	10.323	98	277542	22.603	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.400%
43) trans-1,3-Dichloroprop...	10.579	79	36926	24.588	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.360%
47) 2-Hexanone-d5	10.920	63	26796	59.463	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.920%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	67349	24.641	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.560%
66) 1,2-Dichlorobenzene-d4	13.853	152	89188	22.652	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW020623\
Data File : VW025296.D
Acq On : 06 Feb 2023 15:20
Operator : SY/MD
Sample : 01422-15
Misc : 4.99g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK02

Quant Time: Feb 07 00:20:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020323SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 07 00:17:55 2023
Response via : Initial Calibration

