

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW020720\
 Data File : VW014913.D
 Acq On : 07 Feb 2020 12:08
 Operator : SY/VA
 Sample : L1386-15
 Misc : 4.91G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F3S13

Quant Time: Feb 08 04:39:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 08 03:59:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	815885	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	767022	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	291128	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	243700	18.04	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.16%
7) Chloroethane-d5	2.89	69	244425	22.60	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.40%
10) 1,1-Dichloroethene-d2	4.02	63	378862	15.35	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.40%
20) 2-Butanone-d5	7.07	46	158043	41.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.38%
24) Chloroform-d	7.65	84	558055	24.09	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.36%
26) 1,2-Dichloroethane-d4	8.30	65	316833	24.71	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.84%
29) Benzene-d6	8.27	84	1098606	22.27	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.08%
33) 1,2-Dichloropropane-d6	9.27	67	386953	24.07	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.28%
37) Toluene-d8	10.32	98	941368	21.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.76%
38) trans-1,3-Dichloropropene-	10.58	79	125270	19.74	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.96%
39) 2-Hexanone-d5	10.93	63	107397	36.65	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.30%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	258181	21.33	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.32%
61) 1,2-Dichlorobenzene-d4	13.85	152	281725	24.95	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.80%

Target Compounds					Ovalue
13) Acetone	4.13	43	13020	3.328 ug/L	72
16) Methylene chloride	4.92	84	24729	1.664 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW020720\
Data File : VW014913.D
Acq On : 07 Feb 2020 12:08
Operator : SY/VA
Sample : L1386-15
Misc : 4.91G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F3S13

Quant Time: Feb 08 04:39:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 08 03:59:50 2020
Response via : Initial Calibration

