

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021020\
 Data File : VW014935.D
 Acq On : 10 Feb 2020 11:57
 Operator : SY/VA
 Sample : L1415-11
 Misc : 4.94G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F3S46

Quant Time: Feb 11 00:49:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 11 00:31:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	834229	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	750561	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	272390	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	250621	18.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.60%
7) Chloroethane-d5	2.89	69	250553	22.66	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.64%
10) 1,1-Dichloroethene-d2	4.01	63	421745	16.71	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.84%
20) 2-Butanone-d5	7.07	46	132186	34.10	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.20%
24) Chloroform-d	7.65	84	587172	24.79	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.16%
26) 1,2-Dichloroethane-d4	8.30	65	302436	23.07	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.28%
29) Benzene-d6	8.27	84	1149489	23.82	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.28%
33) 1,2-Dichloropropane-d6	9.27	67	392341	24.94	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.76%
37) Toluene-d8	10.32	98	967846	22.79	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.16%
38) trans-1,3-Dichloropropene-	10.58	79	116069	18.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.76%
39) 2-Hexanone-d5	10.92	63	81290	28.35	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	56.70%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	213677	18.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	72.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	252337	23.89	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.56%

Target Compounds					Ovalue
13) Acetone	4.12	43	18989	4.746 ug/L	96
44) Toluene	10.39	91	23533	0.449 ug/L	93
54) m,p-Xylene	11.83	106	6649	0.318 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021020\
Data File : VW014935.D
Acq On : 10 Feb 2020 11:57
Operator : SY/VA
Sample : L1415-11
Misc : 4.94G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F3S46

Quant Time: Feb 11 00:49:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 11 00:31:29 2020
Response via : Initial Calibration

