

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021020\
 Data File : VW014943.D
 Acq On : 10 Feb 2020 15:26
 Operator : SY/VA
 Sample : L1415-19
 Misc : 5.89G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F3S54

Quant Time: Feb 11 01:20:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 11 00:31:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	829985	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	752073	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	260490	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	223825	16.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.16%
7) Chloroethane-d5	2.89	69	222063	20.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.76%
10) 1,1-Dichloroethene-d2	4.01	63	360776	14.37	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.48%
20) 2-Butanone-d5	7.08	46	142306	36.90	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.80%
24) Chloroform-d	7.65	84	522143	22.16	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.64%
26) 1,2-Dichloroethane-d4	8.31	65	301759	23.14	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.56%
29) Benzene-d6	8.27	84	1031941	21.34	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.36%
33) 1,2-Dichloropropane-d6	9.27	67	372876	23.66	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.64%
37) Toluene-d8	10.32	98	846220	19.88	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.52%
38) trans-1,3-Dichloropropene-	10.57	79	113177	18.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.76%
39) 2-Hexanone-d5	10.92	63	93864	32.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	65.34%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	214055	18.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	72.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	226459	22.42	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.68%

Target Compounds					Ovalue
13) Acetone	4.12	43	17947	4.509 ug/L	84
16) Methylene chloride	4.92	84	25222	1.668 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021020\
Data File : VW014943.D
Acq On : 10 Feb 2020 15:26
Operator : SY/VA
Sample : L1415-19
Misc : 5.89G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F3S54

Quant Time: Feb 11 01:20:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 11 00:31:29 2020
Response via : Initial Calibration

