

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021020\
 Data File : VW014947.D
 Acq On : 10 Feb 2020 17:11
 Operator : SY/VA
 Sample : L1416-04
 Misc : 4.62G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F3S32

Quant Time: Feb 11 02:11:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 11 00:31:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	799183	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	740983	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	264172	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	196243	14.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.32%
7) Chloroethane-d5	2.89	69	195993	18.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.00%
10) 1,1-Dichloroethene-d2	4.01	63	313856	12.98	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.92%
20) 2-Butanone-d5	7.08	46	134144	36.13	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.26%
24) Chloroform-d	7.65	84	470148	20.72	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.88%
26) 1,2-Dichloroethane-d4	8.31	65	269482	21.46	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.84%
29) Benzene-d6	8.27	84	898047	18.85	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.40%
33) 1,2-Dichloropropane-d6	9.27	67	325731	20.98	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.92%
37) Toluene-d8	10.32	98	737156	17.58	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	70.32%
38) trans-1,3-Dichloropropene-	10.57	79	98636	16.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.36%
39) 2-Hexanone-d5	10.92	63	86693	30.63	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.26%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	208029	17.79	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	71.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	206454	20.15	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.60%

Target Compounds					Ovalue
13) Acetone	4.13	43	18415	4.805 ug/L	98
16) Methylene chloride	4.91	84	26727	1.836 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021020\
Data File : VW014947.D
Acq On : 10 Feb 2020 17:11
Operator : SY/VA
Sample : L1416-04
Misc : 4.62G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F3S32

Quant Time: Feb 11 02:11:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 11 00:31:29 2020
Response via : Initial Calibration

