

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW021020\
 Data File : VW014952.D
 Acq On : 10 Feb 2020 19:22
 Operator : SY/VA
 Sample : L1416-09
 Misc : 4.83G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F3S37

Quant Time: Feb 11 02:43:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 11 00:31:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	775132	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	702133	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	237650	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	226162	17.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.48%
7) Chloroethane-d5	2.89	69	226927	22.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.36%
10) 1,1-Dichloroethene-d2	4.01	63	364540	15.55	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.20%
20) 2-Butanone-d5	7.07	46	143497	39.85	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.70%
24) Chloroform-d	7.65	84	560658	25.48	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	101.92%
26) 1,2-Dichloroethane-d4	8.31	65	317548	26.07	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.28%
29) Benzene-d6	8.27	84	1055638	23.38	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.52%
33) 1,2-Dichloropropane-d6	9.27	67	383162	26.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.16%
37) Toluene-d8	10.32	98	873074	21.97	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.88%
38) trans-1,3-Dichloropropene-	10.57	79	114453	19.70	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.80%
39) 2-Hexanone-d5	10.92	63	95595	35.64	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.28%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	245264	22.14	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.56%
61) 1,2-Dichlorobenzene-d4	13.85	152	227752	24.71	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.84%

Target Compounds					Ovalue
13) Acetone	4.12	43	12862	3.460 ug/L	55
16) Methylene chloride	4.92	84	29261	2.072 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW021020\
Data File : VW014952.D
Acq On : 10 Feb 2020 19:22
Operator : SY/VA
Sample : L1416-09
Misc : 4.83G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F3S37

Quant Time: Feb 11 02:43:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM020520S.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 11 00:31:29 2020
Response via : Initial Calibration

