

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021424\
 Data File : VW028000.D
 Acq On : 14 Feb 2024 12:41
 Operator : SY/MD
 Sample : P1442-13
 Misc : 5.62g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9Q6

Quant Time: Feb 15 00:30:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 15 00:29:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	385263	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	344791	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	162827	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	81533	18.069	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.280%
7) Chloroethane-d5	2.905	69	73764	21.553	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.200%
11) 1,1-Dichloroethene-d2	4.021	65	36858	19.090	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.360%
21) 2-Butanone-d5	7.075	46	38560	54.998	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.000%
24) Chloroform-d	7.648	84	200231	22.227	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.920%
26) 1,2-Dichloroethane-d4	8.307	65	97549	23.977	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.920%
32) Benzene-d6	8.276	84	396818	21.511	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.040%
36) 1,2-Dichloropropane-d6	9.270	67	113431	22.678	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.720%
41) Toluene-d8	10.325	98	375058	21.249	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	10.575	79	45448	22.095	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.400%
47) 2-Hexanone-d5	10.922	63	35452	52.109	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.220%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	90173	24.121	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	133119	24.053	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021424\
Data File : VW028000.D
Acq On : 14 Feb 2024 12:41
Operator : SY/MD
Sample : P1442-13
Misc : 5.62g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9Q6

Quant Time: Feb 15 00:30:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 15 00:29:03 2024
Response via : Initial Calibration

