

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021424\
 Data File : VW028010.D
 Acq On : 14 Feb 2024 16:39
 Operator : SY/MD
 Sample : P1469-06
 Misc : 4.81g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9W7

Quant Time: Feb 15 00:32:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 15 00:29:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	472718	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	430779	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	214805	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	82588	14.917	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.680%
7) Chloroethane-d5	2.893	69	77377	18.426	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.720%
11) 1,1-Dichloroethene-d2	4.021	65	38230	16.137	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.560%
21) 2-Butanone-d5	7.075	46	42823	49.778	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.560%
24) Chloroform-d	7.648	84	219038	19.817	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.280%
26) 1,2-Dichloroethane-d4	8.307	65	103952	20.823	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.280%
32) Benzene-d6	8.276	84	433555	18.811	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.240%
36) 1,2-Dichloropropane-d6	9.270	67	123281	19.728	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.920%
41) Toluene-d8	10.319	98	414669	18.804	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.200%
43) trans-1,3-Dichloroprop...	10.575	79	49548	19.280	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.120%
47) 2-Hexanone-d5	10.922	63	42792	50.343	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.680%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	100063	21.423	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.680%
66) 1,2-Dichlorobenzene-d4	13.848	152	143516	19.657	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.640%
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	19120	3.236	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021424\
Data File : VW028010.D
Acq On : 14 Feb 2024 16:39
Operator : SY/MD
Sample : P1469-06
Misc : 4.81g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9W7

Quant Time: Feb 15 00:32:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 15 00:29:03 2024
Response via : Initial Calibration

