

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021624\
 Data File : VW028029.D
 Acq On : 16 Feb 2024 10:42
 Operator : SY/MD
 Sample : P1469-12
 Misc : 5.81g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9X3

Quant Time: Feb 18 23:53:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Feb 18 23:52:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	366460	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	340820	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	171375	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	66771	15.557	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.240%
7) Chloroethane-d5	2.905	69	63463	19.494	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.960%
11) 1,1-Dichloroethene-d2	4.027	65	30817	16.780	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.120%
21) 2-Butanone-d5	7.075	46	40979	61.447	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.900%
24) Chloroform-d	7.648	84	175120	20.437	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.760%
26) 1,2-Dichloroethane-d4	8.307	65	89743	23.190	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.760%
32) Benzene-d6	8.276	84	345603	18.953	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.800%
36) 1,2-Dichloropropane-d6	9.270	67	98622	19.947	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.800%
41) Toluene-d8	10.319	98	331588	19.005	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.040%
43) trans-1,3-Dichloroprop...	10.575	79	42523	20.914	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.640%
47) 2-Hexanone-d5	10.922	63	39307	58.449	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.900%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	92198	24.950	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.800%
66) 1,2-Dichlorobenzene-d4	13.848	152	133012	22.835	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021624\
Data File : VW028029.D
Acq On : 16 Feb 2024 10:42
Operator : SY/MD
Sample : P1469-12
Misc : 5.81g/10mL/MSVOA_W/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9X3

Quant Time: Feb 18 23:53:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Feb 18 23:52:23 2024
Response via : Initial Calibration

