

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021624\
 Data File : VW028031.D
 Acq On : 16 Feb 2024 14:42
 Operator : SY/MD
 Sample : P1469-17
 Misc : 3.46g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9R8

Quant Time: Feb 18 23:53:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Feb 18 23:52:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	347089	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	276931	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	105461	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	71481	17.584	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.320%
7) Chloroethane-d5	2.905	69	67030	21.739	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.960%
11) 1,1-Dichloroethene-d2	4.021	65	32035	18.417	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.680%
21) 2-Butanone-d5	7.081	46	30275	47.930	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.860%
24) Chloroform-d	7.654	84	177796	21.908	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.640%
26) 1,2-Dichloroethane-d4	8.307	65	87225	23.797	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	8.276	84	344904	23.278	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.120%
36) 1,2-Dichloropropane-d6	9.276	67	99587	24.790	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.160%
41) Toluene-d8	10.325	98	310068	21.872	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.480%
43) trans-1,3-Dichloroprop...	10.575	79	36519	22.105	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.400%
47) 2-Hexanone-d5	10.922	63	27972	51.190	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	75432	25.122	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	83401	23.267	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.080%
Target Compounds						
16) Methylene chloride	4.923	84	11088	2.556	ug/L	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021624\
Data File : VW028031.D
Acq On : 16 Feb 2024 14:42
Operator : SY/MD
Sample : P1469-17
Misc : 3.46g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9R8

Quant Time: Feb 18 23:53:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Feb 18 23:52:23 2024
Response via : Initial Calibration

