

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021624\
 Data File : VW028033.D
 Acq On : 16 Feb 2024 15:32
 Operator : SY/MD
 Sample : P1469-16
 Misc : 4.58g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9R7

Quant Time: Feb 18 23:53:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sun Feb 18 23:52:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	413054	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	371267	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	177002	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	76562	15.826	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	63.320%		
7) Chloroethane-d5	2.899	69	71674	19.533	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	78.120%		
11) 1,1-Dichloroethene-d2	4.027	65	35586	17.191	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	68.760%		
21) 2-Butanone-d5	7.081	46	42050	55.940	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	111.880%		
24) Chloroform-d	7.648	84	205039	21.230	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	84.920%		
26) 1,2-Dichloroethane-d4	8.307	65	101977	23.379	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	93.520%		
32) Benzene-d6	8.276	84	408617	20.571	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	82.280%		
36) 1,2-Dichloropropane-d6	9.270	67	115975	21.534	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	86.120%		
41) Toluene-d8	10.318	98	384126	20.211	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	80.840%		
43) trans-1,3-Dichloroprop...	10.575	79	48392	21.849	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	87.400%		
47) 2-Hexanone-d5	10.922	63	41720	56.949	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	113.900%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	101263	25.156	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	100.640%		
66) 1,2-Dichlorobenzene-d4	13.848	152	144112	23.954	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	95.800%		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	12703	2.461	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021624\
Data File : VW028033.D
Acq On : 16 Feb 2024 15:32
Operator : SY/MD
Sample : P1469-16
Misc : 4.58g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9R7

Quant Time: Feb 18 23:53:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Sun Feb 18 23:52:23 2024
Response via : Initial Calibration

