

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022124\
 Data File : VW028048.D
 Acq On : 21 Feb 2024 15:37
 Operator : SY/MD
 Sample : VIBLK596
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK596

Quant Time: Feb 21 23:35:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Feb 21 23:34:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	258450	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	245758	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	123830	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	50697	16.748	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.000%
7) Chloroethane-d5	2.899	69	44830	19.526	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.120%
11) 1,1-Dichloroethene-d2	4.021	65	23304	17.992	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.960%
21) 2-Butanone-d5	7.075	46	22550	47.944	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.880%
24) Chloroform-d	7.648	84	123736	20.475	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.920%
26) 1,2-Dichloroethane-d4	8.307	65	59282	21.720	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.880%
32) Benzene-d6	8.270	84	251101	19.097	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.400%
36) 1,2-Dichloropropane-d6	9.270	67	69796	19.578	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.320%
41) Toluene-d8	10.325	98	236605	18.807	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.240%
43) trans-1,3-Dichloroprop...	10.575	79	28006	19.102	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.400%
47) 2-Hexanone-d5	10.922	63	21315	43.955	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.920%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	56424	21.175	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	88828	21.105	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.400%
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	9035	2.797	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022124\
Data File : VW028048.D
Acq On : 21 Feb 2024 15:37
Operator : SY/MD
Sample : VIBLK596
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK596

Quant Time: Feb 21 23:35:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Feb 21 23:34:58 2024
Response via : Initial Calibration

