

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
 Data File : VW028093.D
 Acq On : 05 Mar 2024 15:14
 Operator : SY/MD
 Sample : P1677-06
 Misc : 13.39g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E06Y0

Quant Time: Mar 05 23:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 05 23:58:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	430739	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	357364	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	142880	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	102492	19.933	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.720%
7) Chloroethane-d5	2.899	69	75947	19.067	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.280%
11) 1,1-Dichloroethene-d2	4.027	65	43353	20.642	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.560%
21) 2-Butanone-d5	7.075	46	36459	49.707	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.420%
24) Chloroform-d	7.648	84	177741	19.482	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.920%
26) 1,2-Dichloroethane-d4	8.307	65	94004	22.140	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.560%
32) Benzene-d6	8.276	84	375777	22.525	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.080%
36) 1,2-Dichloropropane-d6	9.270	67	96515	22.392	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.560%
41) Toluene-d8	10.325	98	345138	21.953	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.800%
43) trans-1,3-Dichloroprop...	10.575	79	44083	23.339	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.360%
47) 2-Hexanone-d5	10.922	63	34520	56.859	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	80953	23.797	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	13.849	152	85344	20.084	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
Data File : VW028093.D
Acq On : 05 Mar 2024 15:14
Operator : SY/MD
Sample : P1677-06
Misc : 13.39g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E06Y0

Quant Time: Mar 05 23:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 05 23:58:07 2024
Response via : Initial Calibration

