

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
 Data File : VW028094.D
 Acq On : 05 Mar 2024 15:38
 Operator : SY/MD
 Sample : P1677-07
 Misc : 8.83g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E06Y1

Quant Time: Mar 06 00:00:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 05 23:58:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	395206	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	316369	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	107880	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	107943	22.881	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.520%
7) Chloroethane-d5	2.899	69	85014	23.263	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.040%
11) 1,1-Dichloroethene-d2	4.033	65	44590	23.139	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	92.560%
21) 2-Butanone-d5	7.075	46	31656	47.039	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.080%
24) Chloroform-d	7.648	84	179272	21.417	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.680%
26) 1,2-Dichloroethane-d4	8.307	65	89650	23.013	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.040%
32) Benzene-d6	8.276	84	370034	25.055	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.200%
36) 1,2-Dichloropropane-d6	9.270	67	95902	25.133	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.520%
41) Toluene-d8	10.325	98	342815	24.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.520%
43) trans-1,3-Dichloroprop...	10.575	79	41561	24.855	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.400%
47) 2-Hexanone-d5	10.922	63	29806	55.456	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.920%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	71950	23.891	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.560%
66) 1,2-Dichlorobenzene-d4	13.848	152	71534	22.296	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.200%
Target Compounds						
					Qvalue	
13) Acetone	4.125	43	6841	7.273	ug/L	73
42) Toluene	10.386	91	12277	0.636	ug/L	96
53) m,p-Xylene	11.837	106	2771	0.323	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
Data File : VW028094.D
Acq On : 05 Mar 2024 15:38
Operator : SY/MD
Sample : P1677-07
Misc : 8.83g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E06Y1

Quant Time: Mar 06 00:00:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 05 23:58:07 2024
Response via : Initial Calibration

