

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
 Data File : VW028095.D
 Acq On : 05 Mar 2024 16:02
 Operator : SY/MD
 Sample : P1677-08
 Misc : 11.83g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E06Y7

Quant Time: Mar 06 00:00:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 05 23:58:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	421789	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	390563	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	197999	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	102164	20.291	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.160%
7) Chloroethane-d5	2.899	69	81263	20.835	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.320%
11) 1,1-Dichloroethene-d2	4.027	65	43590	21.195	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	84.760%
21) 2-Butanone-d5	7.075	46	39705	55.281	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.560%
24) Chloroform-d	7.648	84	178053	19.931	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.720%
26) 1,2-Dichloroethane-d4	8.307	65	97629	23.482	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.920%
32) Benzene-d6	8.276	84	373510	20.486	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.960%
36) 1,2-Dichloropropane-d6	9.270	67	100323	21.297	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.200%
41) Toluene-d8	10.319	98	361650	21.048	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.200%
43) trans-1,3-Dichloroprop...	10.575	79	46263	22.411	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.640%
47) 2-Hexanone-d5	10.922	63	37727	56.859	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	93034	25.023	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.080%
66) 1,2-Dichlorobenzene-d4	13.848	152	129860	22.053	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
Data File : VW028095.D
Acq On : 05 Mar 2024 16:02
Operator : SY/MD
Sample : P1677-08
Misc : 11.83g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E06Y7

Quant Time: Mar 06 00:00:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 05 23:58:07 2024
Response via : Initial Calibration

