

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
 Data File : VW028091.D
 Acq On : 05 Mar 2024 14:26
 Operator : SY/MD
 Sample : P1677-04
 Misc : 5.95g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E06X8

Quant Time: Mar 05 23:59:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 05 23:58:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	441262	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	391450	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	192740	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	109473	20.783	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.120%
7) Chloroethane-d5	2.899	69	85531	20.961	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.840%
11) 1,1-Dichloroethene-d2	4.021	65	48136	22.372	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	89.480%
21) 2-Butanone-d5	7.075	46	30895	41.117	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.240%
24) Chloroform-d	7.648	84	182386	19.515	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.040%
26) 1,2-Dichloroethane-d4	8.307	65	88359	20.314	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.240%
32) Benzene-d6	8.276	84	392764	21.493	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.960%
36) 1,2-Dichloropropane-d6	9.270	67	99968	21.173	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.680%
41) Toluene-d8	10.318	98	378302	21.967	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.880%
43) trans-1,3-Dichloroprop...	10.575	79	43115	20.839	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.360%
47) 2-Hexanone-d5	10.922	63	29046	43.676	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.360%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	75187	20.177	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	116699	20.359	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
Data File : VW028091.D
Acq On : 05 Mar 2024 14:26
Operator : SY/MD
Sample : P1677-04
Misc : 5.95g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E06X8

Quant Time: Mar 05 23:59:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 05 23:58:07 2024
Response via : Initial Calibration

