

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
 Data File : VW028098.D
 Acq On : 05 Mar 2024 17:14
 Operator : SY/MD
 Sample : P1677-11
 Misc : 8.91g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E06Z0

Quant Time: Mar 06 00:00:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 05 23:58:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	442935	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	382953	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	174420	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	107398	20.312	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.240%
7) Chloroethane-d5	2.899	69	85029	20.760	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.040%
11) 1,1-Dichloroethene-d2	4.027	65	44581	20.642	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.560%
21) 2-Butanone-d5	7.075	46	34886	46.253	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.500%
24) Chloroform-d	7.648	84	183937	19.606	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.440%
26) 1,2-Dichloroethane-d4	8.307	65	93686	21.457	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.840%
32) Benzene-d6	8.270	84	391273	21.887	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.560%
36) 1,2-Dichloropropane-d6	9.270	67	101516	21.978	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.920%
41) Toluene-d8	10.319	98	359390	21.332	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.320%
43) trans-1,3-Dichloroprop...	10.575	79	44519	21.995	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.960%
47) 2-Hexanone-d5	10.922	63	32863	50.512	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.020%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	79859	21.906	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.640%
66) 1,2-Dichlorobenzene-d4	13.848	152	108542	20.925	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030524\
Data File : VW028098.D
Acq On : 05 Mar 2024 17:14
Operator : SY/MD
Sample : P1677-11
Misc : 8.91g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E06Z0

Quant Time: Mar 06 00:00:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 05 23:58:07 2024
Response via : Initial Calibration

