

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030722\
 Data File : VW022199.D
 Acq On : 07 Mar 2022 14:28
 Operator : SY/VA
 Sample : N1775-17
 Misc : 6.25g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBSC0

Quant Time: Mar 08 00:24:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 08 00:21:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	169772	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	158085	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	74912	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	32136	14.513	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.040%
7) Chloroethane-d5	2.885	69	30660	17.598	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.400%
11) 1,1-Dichloroethene-d2	4.013	63	46771	13.793	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.160%
21) 2-Butanone-d5	7.080	46	26793	71.964	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	143.920%#
24) Chloroform-d	7.647	84	93811	24.766	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.080%
26) 1,2-Dichloroethane-d4	8.305	65	51136	25.971	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.880%
32) Benzene-d6	8.275	84	165204	22.454	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	9.274	67	52281	25.588	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.360%
41) Toluene-d8	10.323	98	156826	21.441	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.760%
43) trans-1,3-Dichloroprop...	10.573	79	20591	22.729	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.920%
47) 2-Hexanone-d5	10.920	63	23317	69.386	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	138.780%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	52956	29.533	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	118.120%
66) 1,2-Dichlorobenzene-d4	13.847	152	63485	26.385	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030722\
Data File : VW022199.D
Acq On : 07 Mar 2022 14:28
Operator : SY/VA
Sample : N1775-17
Misc : 6.25g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBSC0

Quant Time: Mar 08 00:24:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 08 00:21:43 2022
Response via : Initial Calibration

