

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030722\
 Data File : VW022192.D
 Acq On : 07 Mar 2022 11:33
 Operator : SY/VA
 Sample : N1733-21
 Misc : 7.01g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBRT9

Quant Time: Mar 08 00:22:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 08 00:21:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	169955	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	159401	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	78628	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	27858	12.567	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.280%
7) Chloroethane-d5	2.885	69	25862	14.828	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.320%
11) 1,1-Dichloroethene-d2	4.019	63	41643	12.267	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	49.080%
21) 2-Butanone-d5	7.080	46	18380	49.314	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.620%
24) Chloroform-d	7.647	84	78224	20.629	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.520%
26) 1,2-Dichloroethane-d4	8.311	65	39776	20.180	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.720%
32) Benzene-d6	8.275	84	141119	19.022	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.080%
36) 1,2-Dichloropropane-d6	9.274	67	43003	20.873	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.480%
41) Toluene-d8	10.323	98	128697	17.450	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	69.800%
43) trans-1,3-Dichloroprop...	10.579	79	15721	17.210	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.840%
47) 2-Hexanone-d5	10.920	63	15234	44.959	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.920%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	38466	21.275	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.080%
66) 1,2-Dichlorobenzene-d4	13.853	152	54520	21.588	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.360%
Target Compounds					Qvalue	
13) Acetone	4.123	43	3885	12.762	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030722\
Data File : VW022192.D
Acq On : 07 Mar 2022 11:33
Operator : SY/VA
Sample : N1733-21
Misc : 7.01g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBRT9

Quant Time: Mar 08 00:22:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 08 00:21:43 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 12.762 ug/L
 RT: 4.123 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: VW022192.D
 Acq: 07 Mar 2022 11:33

Instrument :
 MSVOA_W
 ClientSampleId :
 DBRT9

Tgt Ion: 43 Resp: 3885
 Ion Ratio Lower Upper
 43 100
 58 37.2 0.0 69.2

