

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030823\
 Data File : VW025380.D
 Acq On : 08 Mar 2023 12:00
 Operator : SY/MD
 Sample : 01838-02
 Misc : 5.85g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C1169

Quant Time: Mar 09 01:01:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 07 05:10:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	998238	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	887599	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	409062	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	297893	17.948	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.800%
7) Chloroethane-d5	2.905	69	217755	18.233	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.920%
11) 1,1-Dichloroethene-d2	4.027	63	460494	14.365	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.440%
21) 2-Butanone-d5	7.075	46	169133	41.515	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.020%
24) Chloroform-d	7.648	84	550027	19.113	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.440%
26) 1,2-Dichloroethane-d4	8.307	65	324344	20.998	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.000%
32) Benzene-d6	8.276	84	1169183	19.278	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.120%
36) 1,2-Dichloropropane-d6	9.276	67	372965	19.878	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.520%
41) Toluene-d8	10.325	98	1039462	19.439	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.760%
43) trans-1,3-Dichloroprop...	10.575	79	149466	20.006	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.040%
47) 2-Hexanone-d5	10.922	63	105691	39.174	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.340%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	275864	20.069	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.280%
66) 1,2-Dichlorobenzene-d4	13.849	152	315460	20.738	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	82.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030823\
Data File : VW025380.D
Acq On : 08 Mar 2023 12:00
Operator : SY/MD
Sample : 01838-02
Misc : 5.85g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C1169

Quant Time: Mar 09 01:01:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 07 05:10:09 2023
Response via : Initial Calibration

