

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031022\
 Data File : VW022280.D
 Acq On : 10 Mar 2022 17:57
 Operator : SY/VA
 Sample : N1819-03
 Misc : 6.76g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Mar 11 03:11:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 11 03:07:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	202415	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	188853	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	91097	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	70197	26.589	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.360%
7) Chloroethane-d5	2.879	69	55035	26.494	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	105.960%
11) 1,1-Dichloroethene-d2	4.013	63	69805	17.266	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.080%
21) 2-Butanone-d5	7.080	46	20640	46.497	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.000%
24) Chloroform-d	7.647	84	102828	22.769	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.080%
26) 1,2-Dichloroethane-d4	8.384	65	74	0.032	ug/L	0.09
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.120%#
32) Benzene-d6	8.275	84	209103	23.791	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.160%
36) 1,2-Dichloropropane-d6	9.274	67	58692	24.046	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.200%
41) Toluene-d8	10.323	98	199899	22.877	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.520%
43) trans-1,3-Dichloroprop...	10.573	79	24032	22.206	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.840%
47) 2-Hexanone-d5	10.921	63	17014	42.381	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.760%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	46946	21.916	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	72230	24.686	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.760%
Target Compounds						Qvalue
37) 1,2-Dichloropropane	9.268	63	6906	3.242	ug/L	# 85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031022\
Data File : VW022280.D
Acq On : 10 Mar 2022 17:57
Operator : SY/VA
Sample : N1819-03
Misc : 6.76g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Mar 11 03:11:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 11 03:07:48 2022
Response via : Initial Calibration





#37

1,2-Dichloropropane

Concen: 3.242 ug/L

RT: 9.268 min Scan# 11

Delta R.T. -0.098 min


Lab File: VW022280.D

Acq: 10 Mar 2022 17:57

Instrument :

MSVOA_W

ClientSampleId :



Tgt Ion: 63 Resp: 6906

Ion Ratio Lower Upper

63 100

112 0.0 4.1 6.1#

