

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031022\
 Data File : VW022271.D
 Acq On : 10 Mar 2022 14:16
 Operator : SY/VA
 Sample : N1818-16
 Misc : 6.62g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBSG4

Quant Time: Mar 11 03:10:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 11 03:07:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	199659	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	185448	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	90709	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	62463	23.986	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.960%
7) Chloroethane-d5	2.879	69	44544	21.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.960%
11) 1,1-Dichloroethene-d2	4.007	63	66156	16.589	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.360%
21) 2-Butanone-d5	7.074	46	22399	51.156	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.320%
24) Chloroform-d	7.647	84	99809	22.405	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.640%
26) 1,2-Dichloroethane-d4	8.305	65	54106	23.366	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.480%
32) Benzene-d6	8.275	84	202376	23.448	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.800%
36) 1,2-Dichloropropane-d6	9.274	67	55132	23.002	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.000%
41) Toluene-d8	10.323	98	196093	22.853	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.400%
43) trans-1,3-Dichloroprop...	10.573	79	24189	22.761	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.040%
47) 2-Hexanone-d5	10.921	63	18338	46.518	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.040%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	47114	22.398	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	68957	23.668	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031022\
Data File : VW022271.D
Acq On : 10 Mar 2022 14:16
Operator : SY/VA
Sample : N1818-16
Misc : 6.62g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBSG4

Quant Time: Mar 11 03:10:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 11 03:07:48 2022
Response via : Initial Calibration

