

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031723\
 Data File : VW025408.D
 Acq On : 17 Mar 2023 10:26
 Operator : SY/MD
 Sample : VW0317SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK471

Quant Time: Mar 17 21:40:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 17 21:39:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	728642	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	640273	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	296196	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	256311	21.156	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.640%
7) Chloroethane-d5	2.899	69	199653	22.903	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	4.021	63	417899	17.859	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.440%
21) 2-Butanone-d5	7.075	46	154419	51.927	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.860%
24) Chloroform-d	7.648	84	512430	24.395	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.560%
26) 1,2-Dichloroethane-d4	8.307	65	283318	25.129	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.520%
32) Benzene-d6	8.276	84	1078198	24.645	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.560%
36) 1,2-Dichloropropane-d6	9.276	67	337520	24.938	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.760%
41) Toluene-d8	10.325	98	943558	24.462	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.840%
43) trans-1,3-Dichloroprop...	10.575	79	126899	23.546	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.200%
47) 2-Hexanone-d5	10.922	63	97079	49.881	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.760%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	247951	25.006	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.040%
66) 1,2-Dichlorobenzene-d4	13.849	152	288963	26.234	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031723\
Data File : VW025408.D
Acq On : 17 Mar 2023 10:26
Operator : SY/MD
Sample : VW0317SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK471

Quant Time: Mar 17 21:40:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 17 21:39:25 2023
Response via : Initial Calibration

