

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032123\
 Data File : VW025473.D
 Acq On : 21 Mar 2023 17:04
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK425

Quant Time: Mar 22 01:35:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Mar 22 01:30:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	730751	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	654856	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	303565	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	223647	18.407	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.640%
7) Chloroethane-d5	2.893	69	175757	20.104	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.400%
11) 1,1-Dichloroethene-d2	4.027	63	367567	15.663	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.640%
21) 2-Butanone-d5	7.075	46	159938	53.628	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.260%
24) Chloroform-d	7.728	84	571	0.027	ug/L	0.08
Spiked Amount	25.000	Range	40 - 150	Recovery	=	0.120%#
26) 1,2-Dichloroethane-d4	8.185	65	185	0.016	ug/L	-0.12
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.080%#
32) Benzene-d6	8.270	84	1015482	22.694	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.760%
36) 1,2-Dichloropropane-d6	9.270	67	326798	23.608	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.440%
41) Toluene-d8	10.319	98	874527	22.167	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.680%
43) trans-1,3-Dichloroprop...	10.575	79	120908	21.935	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.720%
47) 2-Hexanone-d5	10.922	63	101111	50.796	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.600%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	249644	24.616	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	282550	25.029	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032123\
Data File : VW025473.D
Acq On : 21 Mar 2023 17:04
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK425

Quant Time: Mar 22 01:35:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Mar 22 01:30:00 2023
Response via : Initial Calibration

